

QCS Laboratory Program Update

Laboratory Advisory Committee Meeting
September 9, 2019

Steven Sievert

Manager, Quality Certification Services Inc.

Technical Director, National DHIA

Chair, ICAR Subcommittee for Recording and Sampling Devices

Housekeeping

Auditing Schedules

Current list of labs with respective centering months included

- 22 labs in even years
- 21 labs in odd years
- Lab audits conducted within 60 days of centering month based on auditor's schedule

Schedule modifications

- QCS works with lab auditor to determine the most efficient use of resources – time, travel

Availability of Samples During On-Site Audit

Laboratory **MUST** have samples to run the day of the on-site audit. If there are no samples available, the on-site audit will be terminated and will have to be rescheduled.

Laboratory is responsible for all costs (time and travel) associated with the subsequent audit.

Will negatively affect your certification status (i.e. Provisional).

Note that the certification expiration date cannot be extended and the auditor's schedule may push subsequent audit date past the existing expiration date. The net result is decertification of the laboratory until the on-site audit can be completed. Decertified laboratories may not send data to the CDCB.

Non-Compliant Issues from Previous On-Site Audit

It is normal that certain noncompliant items identified during the course of the on-site audit are designated with a completion timeline of ‘by the next audit’

If a laboratory fails to address these noncompliant items by the subsequent audit, the laboratory will have its certification status changed to ‘Conditional.’

May bypass the ‘Conditional’ status if additional serious noncompliant issues are identified during the course of the subsequent audit.

The auditor will recommend to QCS a time-frame for completion that will not exceed six (6) months.

Failure to address these items within the time-frame designated will result in the laboratory certification status to be changed to ‘Provisional.’ If a laboratory continues to fail to address the noted noncompliant issues, the laboratory may be decertified.

Proposed Changes to

Auditing Procedures for Laboratories

Addition of Full Spectral Fat Calibration to Guidelines

Proposed by Dairy One Cooperative Inc.

- **Either 'B' wavelength or full spectral calibration**
- **Supporting documentation circulated to all labs on August 31, 2018**

Changes to text of Guidelines include:

Butterfat ~~Analysis and "B" Filter~~

All IR analyzers must use either a "B" wavelength or full spectral calibration.

New Instrument Approval Protocol

1. As outlined in the *General Auditing Guidelines*, the new instrument(s) must be reported to QC Program Manager and subsequently enrolled in the monthly Samples Unknown program. For each new instrument, the following information should be provided:
 - a. Manufacturer,
 - b. Model,
 - c. Condition (new, used, refurbished),
 - d. Serial number,
 - e. Components to be analyzed (fat, protein, SCC, MUN, other),
 - e.f. Fat analysis (if applicable) using "B" wavelength or full spectra,
 - f.g. Instrument(s) to be replaced/taken out of service (where applicable).


Quantitative PT Programs

Samples Unknown

February 2019 Fat PT Results


- Disagreement between reference chemistry and instrument results
- Different models of instruments provided clustered results
- Most likely related to fatty acid profile of sample sets (calibration/PT)

PT Sample Set Comparison -/ Butterfat



August 2019 Fat PT Results


PT Sample Set Comparison -/ Butterfat



August 2019 True Protein PT Results

- Slight bias (0.012) on protein results
- Bias most likely introduced by calibration samples prior to PT sample week – most laboratories did not notice the shift


PT Sample Set Comparison - / True Protein



August 2019 MUN PT Results


- Significant differences in results by instrument type however the majority of instruments are within conformance standards.
- Opportunity to work with manufacturers if labs agree to share data.

PT Sample Set Comparison -/ Milk Urea Nitrogen



August 2019 SCC PT Results

PT Sample Set Comparison -/ SCC



Adding a New Instrument


Notify QCS Program Manager of new instrument:

- **Make, Model and In-Service Date**
- **Components to be analyzed**
- **Instrument to be taken off-line (if applicable)**

Laboratory adds instrument on PT platform.

Documentation Required

- **Manufacturer training is required and subsequent documentation sent to the QCS Program Manager**
- **Analyze one set of 'special' unknowns with results sent to QC Program Manager and Paul Sauvé.**
- **Perform appropriate and routine QC checks with calibration checks, hourlies and dailies for the first three weeks of operation with results sent to Steven & Paul.**



Managing Equipment at Your Lab

Manage Equipment

Equipment				
Purpose	Name	Active	Instrument Type	Supported Properties
Routine	Station A	Yes		Retire
Routine	FOSS 6000	Yes	FOSS 6000	Butterfat, Protein, Milk Urea Nitrogen
Routine	FOSS 5000	Yes	FOSS 5000	SCC
Routine	Station B	Yes		Retire
Routine	FOSS 6000	Yes	FOSS 6000	Butterfat, Protein, Milk Urea Nitrogen
Routine	FOSS FC	Yes	FOSS FC	SCC
Routine	Station C	Yes		Retire
Routine	FOSS FT+	Yes	FOSS FT+	Butterfat, Protein, Milk Urea Nitrogen
Routine	FOSS 5000	Yes	FOSS 5000	SCC
Routine	Station D	Yes		Retire
Routine	FOSS FT+	Yes	FOSS FT+	Butterfat, Protein, Milk Urea Nitrogen
Routine	FOSS FC	Yes	FOSS FC	SCC
Routine	Station E	Yes		Retire
Routine	FOSS FT+	Yes	FOSS FT+	Butterfat, Protein, Milk Urea Nitrogen
Routine	FOSS FC	Yes	FOSS FC	SCC
Reference	REF CHEM	Yes		Retire
Routine	Old Station 1	No		
Routine	FOSS 4000	No	FOSS 4000	Butterfat, Protein, Milk Urea Nitrogen
Routine	Old Station 2	No		
Routine	FOSS 4000	No	FOSS 4000	Butterfat, Protein, Milk Urea Nitrogen
Add Station				

Retiring Stations and Instruments

Quality Certification Services Inc. Demo Laboratory Setup

- Manage Equipment
- Manage Users

Demo Laboratory Dashboard

Quality Certification Services Inc. Demo Laboratory

qcs.test.analytictrust.com says:
Do you want to retire this equipment?

OK Cancel

Equipment				
Purpose	Name	Active	Instrument Type	Supported Properties
Routine	Station A	Yes		Retire
Routine	FOSS 6000	Yes	FOSS 6000	Butterfat, Protein, Milk Urea Nitrogen
Routine	FOSS 5000	Yes	FOSS 5000	SCC

Quality Certification Services Inc. Demo Laboratory

qcs.test.analytictrust.com says:
Do you want to retire this instrument?

OK Cancel

Demo Laboratory / Station

[Edit] [Back to Equipment List]

Name Station A

Instruments			
Active	Name	Type	Supported Properties
Yes	FOSS 5000	FOSS 5000	SCC Retire
Yes	FOSS 6000	FOSS 6000	Butterfat, Protein, Milk Urea Nitrogen Retire

Add Instrument

Adding a New Station

Adding a New Instrument

Create New Station

Station Details

Name*

Create and Close

Station A - Create New Instrument

[\[Back to Station \]](#)

Instrument details

Name*

Serial Number*

Instrument Type*

-- Please select --

- FOSS 300
- FOSS 303
- FOSS 360
- FOSS 400
- FOSS 4000
- FOSS 4000/5000
- FOSS 4000/FOSS 5000
- FOSS 5000
- FOSS 6000
- FOSS 6000/400
- FOSS 6000/5000
- FOSS 6000/FC
- FOSS 605
- FOSS 7**
- FOSS FC
- FOSS FT
- FOSS FT+
- FOSS FT+/5000
- FOSS FT+/FC
- Skalar

Station A - 237

Start Date 8/6/2017

Due Date 8/26/2017

Instruments

FOSS 6000

Serial number:

Properties: Milk Urea Nitrogen, Protein, Butterfat

Take Offline

FOSS 5000

Serial number:


Properties: SCC

Take Offline

Marking Instruments Offline During a Monthly Trial

Must enter the number of samples run on the instrument during the previous month

24	Value	Value	Value	Value
Totals	1 - 12: - 13 - 24: -	1 - 12: - 13 - 24: -	1 - 12: - 13 - 24: -	1 - 12: - 13 - 24: -
Total Milk Samples Run During Previous Calendar Month for this Instrument*		<input type="text"/>		
Monthly Instrument Notes		<input type="text"/>		
<input type="button" value="Save Measurements"/>				



Qualitative PT Programs

Milk ELISA PT for MAP & PAG

QCS Qualitative PT Platform

Designed for multiple qualitative tests

- **MAP**
- **PAG**
- **Others (BLV, BVD, etc.)**
- **Milk or Serum based tests**

Multiple manufacturers in one PT

- **Specific controls for each test kit**
- **Administrative overview sorted by manufacturer**
- **Manufacturer only sees results of those technicians using their test kit**

Admin View of PT Results

- View of all results to identify trends, sample validity and identify laboratories for follow-up
- Helped gather information related to challenges in quantifying the suspect range
- Internal report only – not shared with individual labs at this time

	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	+	?	-	(+)	(?)	(-)	□
	-	-	+	-	-	?	+	+	?	-	+	?	-	+	+	-	?	-	+	+	100.0%	-	100.0%	33.3%	-	-	-
	-	+	-	-	-	(+)	+	+	(+)	-	+	(+)	-	+	+	-	(+)	-	+	+	100.0%	75.0%	100.0%	8.3%	-	-	-
	-	+	-	-	-	?	+	+	?	-	+	?	-	+	+	-	(+)	-	+	+	100.0%	50.0%	100.0%	-	-	16.7%	-
	-	+	-	-	-	(-)	+	+	?	-	+	?	-	+	+	-	(-)	-	+	+	100.0%	50.0%	100.0%	16.7%	-	-	-
	-	+	-	-	-	?	+	+	(+)	-	+	?	-	+	+	-	(+)	-	+	+	100.0%	50.0%	100.0%	16.7%	-	-	-
	-	+	-	-	-	?	+	+	(+)	-	+	?	-	+	+	-	(+)	-	+	+	100.0%	50.0%	100.0%	16.7%	12.5%	-	-
	-	+	-	-	-	?	+	+	(+)	-	+	(+)	(?)	+	+	(?)	?	-	+	+	100.0%	50.0%	75.0%	16.7%	12.5%	-	-
	-	+	-	-	-	?	+	+	?	-	+	?	-	+	+	-	?	-	+	+	100.0%	100.0%	100.0%	-	-	-	-
	-	+	-	-	-	?	+	+	?	-	+	?	-	+	+	-	?	-	+	+	100.0%	-	-	-	-	-	-
	-	+	-	-	-	?	+	+	?	-	+	?	-	+	+	-	?	-	+	+	100.0%	-	-	-	-	-	-
	-	+	-	-	-	?	+	+	?	-	+	?	-	+	+	-	?	-	+	+	100.0%	-	-	-	-	-	-
	-	+	-	-	-	?	+	+	?	-	+	?	-	+	+	-	?	-	+	+	100.0%	75.0%	100.0%	-	-	8.3%	-
	-	+	-	-	-	(+)	+	+	(+)	-	+	(+)	-	+	+	-	(+)	-	+	+	100.0%	-	100.0%	33.3%	-	-	-
	-	+	-	-	-	(+)	+	+	(+)	-	+	(+)	-	+	+	-	(+)	-	+	+	100.0%	-	100.0%	33.3%	-	-	-
	-	+	-	-	-	?	+	+	?	-	+	?	-	+	+	-	?	-	+	+	100.0%	100.0%	100.0%	-	-	-	-
	-	+	-	-	-	(+)	+	+	(+)	-	+	?	-	+	+	-	(+)	-	+	+	100.0%	25.0%	100.0%	25.0%	-	-	-
	-	+	-	-	-	?	+	+	?	-	+	?	-	+	+	-	?	-	+	+	100.0%	-	-	-	-	-	-
	-	+	-	-	-	?	+	+	?	-	+	?	-	+	+	-	?	-	+	+	100.0%	75.0%	100.0%	-	-	8.3%	-
	-	+	-	-	-	?	+	+	?	-	+	?	-	+	+	-	?	-	+	+	100.0%	100.0%	100.0%	-	-	-	-
	-	+	-	-	-	?	+	+	(+)	-	+	?	-	+	+	-	(+)	-	+	+	100.0%	50.0%	100.0%	16.7%	-	-	-
	-	+	-	-	-	?	+	+	(+)	-	+	(+)	-	+	+	-	(+)	-	+	+	100.0%	25.0%	100.0%	25.0%	-	-	-

ELISA PT Report


- Removed overall percentage – there is no ‘passing score’
- Shows correct & false positives, negatives, & suspects
- Incorrect suspects do NOT affect certification status
- QC is an on-going process that involves QMS in lab, PT testing & manufacturer support

	Positive	Suspect	Negative
Correct	100.0%	100.0%	100.0%
False	-	-	-

Current PT Sample Set Results

Sample	Reference	Site	Difference	Result	Conclusion
1	0.062	0.055	0.007	- -	Correct Negative
2	1.770	1.927	-0.157	+ +	Correct Positive
3	0.443	0.427	0.016	+ -	Correct Positive
4	1.543	1.666	-0.123	+ -	Correct Positive
5	0.420	0.378	0.042	+ +	Correct Positive
6	0.063	0.060	0.003	- -	Correct Negative
7	1.546	1.817	-0.271	+ +	Correct Positive
8	1.819	2.015	-0.196	+ -	Correct Positive
9	0.068	0.054	0.014	- -	Correct Negative
10	0.063	0.057	0.006	- -	Correct Negative
11	1.342	1.733	-0.391	+ +	Correct Positive
12	0.063	0.062	0.001	- -	Correct Negative
13	0.070	0.057	0.013	- -	Correct Negative
14	0.066	0.057	0.009	- -	Correct Negative
15	1.522	1.746	-0.224	+ +	Correct Positive
16	1.747	2.024	-0.277	+ -	Correct Positive
17	0.295	0.347	-0.052	? ?	Correct Suspect
18	1.775	1.926	-0.151	+ +	Correct Positive
19	0.068	0.054	0.014	- -	Correct Negative
20	0.457	0.382	0.075	+ -	Correct Positive

Regression Analysis



Historical Performance

Conclusion	Trial 200 - Johne's	Trial 199 - Johne's	Trial 198 - Johne's	Trial 197 - Johne's	Last 3 Months	Last 6 Months	Last 12 Months
Correct Positive	100.0%	100.0%	-	100.0%	100.0%	100.0%	100.0%
Correct Suspect	100.0%	100.0%	-	-	100.0%	68.7%	68.7%
Correct Negative	100.0%	100.0%	-	100.0%	100.0%	100.0%	100.0%
False Positive	-	-	-	28.6%	-	9.5%	9.5%
False Suspect	-	-	-	-	-	-	-
False Negative	-	-	-	-	-	-	-
Missing	-	-	100.0%	-	-	-	-
Bad measurement	-	-	-	-	-	-	-

This report may not be reproduced, except in full, without the written permission of Quality Certification Services, Inc. This report is final and only applies to the item(s) tested in the time period specified and complies with ISO 17025:2017.

Steven Sievert
Manager, Quality Certification Services

Calculation of S/P Ratio

$$\text{S/P} = \frac{(\text{Sample OD} - \text{Mean NC})}{(\text{Mean PC} - \text{Mean NC})}$$

IDEXX MAP

S/P Calculation and Interpretation

Qualitative Call	Milk	Serum/Plasma
Negative	$\text{S/P} \leq 0.20$	$\text{S/P} \leq 0.45$
Suspect	$0.20 > \text{S/P} < 0.30$	$0.45 > \text{S/P} < 0.55$
Positive	$\text{S/P} \geq 0.30$	$\text{S/P} \geq 0.55$

**(Old suspect range was 0.30 to 0.40 –
changed in 2015)**

ELISA MAP Control Results

- Actual negative & positive controls as reported in PT program
- Positive Controls ranged from 0.490 to 1.935
- A suspect sample with OD of 0.250 will result in varying S/P ratios and qualitative calls
- What is normal variation and what is signal for action?

Test Kit	Negative 1	Negative 2	Positive 1	Positive 2	0.250 OD	Call
Ref	0.031	0.032	1.082	1.059	0.210	SUS
7067	0.037	0.038	0.882	0.833	0.259	SUS
7097	0.049	0.040	1.378	1.311	0.158	NEG
7067	0.043	0.042	0.922	0.903	0.239	SUS
7067	0.045	0.045	1.000	0.888	0.228	SUS
7067	0.038	0.039	1.181	1.196	0.184	NEG
7067	0.049	0.049	0.689	0.677	0.317	POS
7097	0.049	0.040	1.378	1.311	0.158	NEG
7097	0.045	0.041	1.687	1.634	0.128	NEG
7067	0.044	0.045	0.984	0.911	0.228	SUS
7067	0.039	0.038	1.425	1.444	0.152	NEG
7067	0.042	0.042	1.224	1.238	0.175	NEG
7067	0.037	0.039	0.832	0.845	0.265	SUS
7067	0.040	0.040	1.309	1.301	0.166	NEG
7067	0.059	0.049	1.293	1.322	0.156	NEG
7097	0.044	0.044	1.305	1.428	0.156	NEG
7097	0.051	0.050	0.583	0.591	0.372	POS
7097	0.050	0.055	0.639	0.676	0.375	POS
7067	0.022	0.024	0.841	0.855	0.275	SUS
7067	0.048	0.049	1.952	1.917	0.107	NEG
7097	0.042	0.042	1.013	1.041	0.211	SUS
7067	0.042	0.044	0.746	0.721	0.300	POS
7067	0.042	0.047	0.487	0.492	0.462	POS
7097	0.040	0.086	0.852	0.828	0.241	SUS
7097	0.040	0.042	0.948	0.857	0.243	SUS

August 2019

ELISA MAP


Control Results

- Initial Reference Lab Positive
Controls ranged averaged 0.705
- All suspects received call of positive based on S/P calculation
- Retest by reference lab and updated reports coming
- Causes?
 - Pipettor?
 - TMB?
 - Plate Washer?
 - Control Plate for Plate Reader?
 - Procedural Drift?

	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20
	-	-	+	+	+	+	+	-	+	-	-	+	+	-	+	-	-	+	+	+
	-	-	+	+	+	+	+	-	+	-	-	(?)	+	-	+	-	-	+	(?)	(?)
	-	-	+	+	+	+	+	-	+	-	-	+	+	-	+	-	-	+	+	+
	-	-	+	+	+	+	+	-	+	-	-	(?)	+	-	+	-	-	(?)	(?)	(?)
	-	-	+	+	+	+	+	-	+	-	-	(?)	+	-	+	-	-	(?)	(-)	(-)
	-	-	+	+	+	+	+	-	+	-	-	(?)	+	-	+	-	-	(?)	(-)	(-)
	-	-	+	+	+	+	+	-	+	-	-	(?)	+	-	+	-	-	(?)	(?)	(?)
	-	-	+	+	+	+	+	-	+	-	-	(?)	+	-	+	-	-	(?)	(?)	(?)
	-	-	+	+	+	+	+	-	+	-	-	(?)	+	-	+	-	-	(?)	(?)	(?)
	-	-	+	+	+	+	+	-	+	-	-	(?)	+	-	+	-	-	(?)	(?)	(?)
	-	-	+	+	+	+	+	-	+	-	-	(?)	+	-	+	-	-	(?)	(?)	(?)
	-	-	+	+	+	+	+	-	+	-	-	(?)	+	-	+	-	-	(?)	(?)	(-)
	-	-	+	+	+	+	+	-	+	-	-	(?)	+	-	+	-	-	(?)	(-)	(-)

MAP ELISA

Drop in Percentage of Positive Calls



- Change noticed in October 2017 by Minnesota DHIA
- Thoughts and Discussion?

Milk Pregnancy PT Launch

Final details in Fall 2019

- Number of samples - 12
- Two each - pregnant, recheck, open
- Frequency – bimonthly (alternate with MAP PT test)
- All technicians performing assay participate in each test for PAG & MAP
- Multiple manufacturers supported

Steps to be taken

- 3-4 labs to test PT trial sensitivity – volunteers?
- Enrollment form from QCS
- Manufacturer support
- Circle of feedback and support to be finalized

ELISA Milk Pregnancy

Focus on 'Rechecks'

MILK ELISA PERCENT RECHECKS IN 2019



- Data from one lab – 19,916 ELISA PAG tests from January 2, 2019 through August 30, 2019
- Increased percent of rechecks reported by multiple labs – many over 6% recently
- Thoughts and Discussion?